

# Measurement the Chemical Composition and Topography of Copper Nickel Titanium Arch Wires at Dry and Wet conditions (An *In-Vitro* Study)

Saba Ammar Mohammed<sup>1</sup>, Abeer Basim Mahmood<sup>2</sup>, Mohammed Nahidh<sup>2</sup>

<sup>1</sup>Master Student, <sup>2</sup>Assit. Prof. Department of Orthodontics, College of Dentistry, University of Baghdad, Baghdad/  
Iraq

## Abstract

**Background:** The characteristics surface, such as, topography, and rigidity of orthodontic arch wires are essential determining factor of the effectiveness, of arch wire-guided tooth movement. They also affect, in the potential of corrosion, and the aesthetics, of orthodontic components. This study was considered to evaluate the chemical composition and topography of Cu-NiTi arch wires from different companies in, dry and, wet condition( acidic and, neutral artificial saliva). **Materials and methods** Three, types of CuNiTi orthodontic, archwires from, different brands were, used: CuNiTi (Ormco brand),, CuNiTi (IOS brand) and, CuNiTi (Orthotechnology brand),. They were rectangular, (0.018\*0.025 inch) in cross, section and cut into pieces, of 20 mm in length., 30 pieces, from each, company were separated into 3 groups; one group was left dry condition, and other, 2 groups, ten pieces, for each group were submerged in, artificial saliva, (pH=6.75 ± 0.015) and, (pH=3.5 ± 0.015) at 37°C , for 2 months. The SEM and EDS was used to assess surface analysis of wholly samples., **Results:** After soaking period, CuNiTi wires from orthotechnology brand, exhibited the highest change in chemical composition and topography among the study groups. Instead, the least change in chemical composition and topography surfaces were CuNiTi wires from Ormco brand, when compared to, analogous arch wires. **Conclusion:** The CuNiTi arch wires from Ormco brand were the greatest and the most appropriate alignment in orthodontic archwires in term of the least changes in surface of Cu-NiTi initially and during the course of this study period.

**Keywords:** Copper Nickel Titanium arch wires, scanning-electron microscopy and Energy dis persive x-ray spectrophotometer .

## Introduction

An appliance continued fixed on teeth within, attachment by, which the request of force occurred by arch wires or auxiliaries is called fix orthodontic appliance <sup>(1)</sup>. The one of most main portions of fixed appliance is arch wire, the perfect one should have capacity to move the tooth with a, continuous ,light force applied to it <sup>(2)</sup>. Nickel titanium- Copper chromium Alloy, (CuNiTi): This alloy was advanced by Dr. Rohit Sachdeva and Miyasaki, in

1994. It composed of (49.87%), nickel, (5.64%), Copper, (42.99%) titanium ,and (0.5%) chromium''''<sup>(3)</sup>. To increase the thermal-reactive assets of the archwire and increase the transformation of the temperature to above the oral cavity and to recompense for this, copper additional to the alloy, the chromium is added to reduction the transformation of the temperature. Originally, '' Copper NiTi was produced with four altered austenitic transformation temperatures covering both thermoelastic archwires and pseudoelastic ''<sup>(4)</sup>. The cobalt-chromium, stainless-steel,, and titanium alloys were, used in the orthodontic, appliances depend on the formation of passive oxide surface films to corrosion resist. These layers of protective are not, reliable. They are prone to both chemical and mechanical degradation.

---

**Corresponding author:**

**Mohammed Nahidh**

ali.mario28@yahoo.com

Even, without destruction, oxide films regularly dissolve slowly only to alteration as metal surface is exposed to oxygen from the surrounding medium or from air [5].

The purpose of the present study was to analyze chemical composition and topography of copper nickel titanium archwires at dry and neutral, acidic or artificial saliva.

## Materials and Methods

### Formation of Artificial Saliva

The mechanisms of artificial saliva ‘‘ 0.7g NaCl , 1.2g KCl , 0.26g Na<sub>2</sub>HPO<sub>4</sub> , 0.2g K<sub>2</sub>HPO<sub>4</sub> , 1.5g NaHCO<sub>3</sub> , 0.33g KSCN , 0.13g, urea and 1000,ml deionized water’’ .’’ NaOH and Lactic acid were used to alter ‘‘the pH of artificial saliva using the pH-meter’’ and kept in, 37°C using incubator using filter paper after filtering to get clear of any unsolvable impurities and salts. The, pH of artificial saliva, was attuned to 6.75±0.015 and 3.5±0.015 using a pH meter (Jenway,, Cyprus and model 3320, equivalent to the pH of human salivary’’ (6-8).

### The Samples

The samples, were made of (3) types of maxillary archwires rectangular in cross-section ((0.018\*0.025

inch)):


1. CuNiTi, archwires (Ormco corp., Glendora, CA, USA),

2. Copper Nickel, Titanium archwire, ((Ortho Technology®, USA, TruFlex™, full form))

And 3. CuNiTi archwires (IOS, International Orthodontic Services, Stafford, USA ).

The straight parts of the arch wire were cut into 20 mm in length (9). Totally was 90 pieces of arch wire, 30, pieces from each type, of arch wires, were divided in, the groups that 10 parts from each, company retain in a dry, condition ‘‘as a control group’’, although the other, 20 parts, were submerged, in artificial saliva, for 2 month, ten pieces, from each, groups were submerged, in pH, of artificial, of saliva 6.75 and , pH = 3.5.

Afterward, samples, was employed, in glass containers, separated and held in from its, ends held by, dental floss .Then, , added of artificial, saliva so that, sample, was submerged, completely. (Figure 1).



Figure(1): samples in the artificial saliva.

Then, the samples, were saved in an incubator, at 37 °C in (Fisher scientific) for two months. Then, artificial saliva was changed regularly every seven days with a fresh solution to evade its saturation with the products of the degradation (10,11). After the 2 months the samples were washed by distilled water. Then, left to dry and preserved in petri dishes.

### Specimens

then, it desires to use minor slides. With the determination of use scanning electron microscope and energy dispersive x-ray spectrophotometer for analysis the sample. The cut of slide by a diamond cutting instrument. Then, the sample fixed on a slide.


The studies of orthodontic materials in artificial salivas on the corrosion, resistance are very popular. In

the current work, the changes of the surfaces of CuNiTi alloy of the orthodontic wires in neutral artificial saliva pH=6.75 and acidic artificial saliva pH=3.5 after 2 months are examined by the Scanning Electron Microscopy and Energy dispersive x-ray spectrophotometer EDS methods.

### Results

#### Surface micromorphology of the Copper Nickel Titanium in dry and neutral, acidic artificial saliva

Surface micromorphology of the Copper Nickel Titanium was evaluated before and after immersion in neutral and acidic of artificial saliva, visualized at 2000X magnification using SEM and the surface topography was assessed using EDS, as shown in (figures: 2) in dry condition.



**Figure (2): Surface micromorphology of CuNiTi arch wires at 2000X magnification using SEM. A) represents the surface of ormco brand at dry condition ; b) represents the surface of IOS brand at dry condition ; c) represents the surface of orthotechnology brand at dry condition.**

It was found that after CuNiTi archwires immersion in artificial saliva exaggerates surface changes of the wires, this was represented by appearance of numerous pits with different depth and size, as being reflected by the increase in number and height of beaks. Concomitant with immersion in artificial saliva for 2 months.

So that the control groups (dry condition) exhibited increased number of pits with shallow depth. It is possible, that the pits represented that roughness and

artificial saliva enhanced the depth and number of these pits. Indeed, with orthotechnology and IOS brands, major changes occurred representative by presence of great fissures connecting pits and area of surface break. Nevertheless, with Ormco brand, the pits were more distributed regularly through the surface with the smallest changes, occurred compared to other brands.



The SEM examination, formed oxide-based particles seem in dark - gray contrast (Fig.3). It is obviously seen concentration of oxygen is higher, than in the matrix. This means that, Ti-based, oxide is,formed as result of tough interaction stuck between the solution and metal. The dissolution,of nickel, from the matrix, into the solution, is very,high. All effects,can be,traced by ,considering the ,concentration differences in the EDS,analyses. The EDS mapping shows to be very beneficial in understanding ,the elemental, distribution for,given materials. Nickel, and titanium,are the major elements on surface and oxide-based, particles can be simply separated from,the matrix because of their oxygen ,level.

The EDS ,studies show to be very beneficial in determining components,formed on the,surface of the,metal as result of corrosion,in the artificial saliva ,after a period ,of time.

The results OF EDS shown in (Table 1) indicate that a clear reduction in the amounts of Ni and Ti because of the dissolution,of the metals.From the,clinical point,of view,,the final result of orthodontic,treatment may,be compromised,by corrosion, and metal ions appearing as corrosion,products (Ni<sup>2+</sup> ) may,result in,symptoms of,toxicity andallergic reactions <sup>(12)</sup>. The oxide,film behaves,as a protective,layer during,corrosion in,aggressive environments <sup>(13)</sup>.

“ Titanium exists in the solution as TiO<sub>2</sub> having a more stable form”. Ni element reduction in composition,of wires and,Ti increase,that because could be,attributed to,the fact that addition of small quantity,of Cu to,the structure of archwires reduced the,reactivity of Ti and increased resistance to,corrosion and,biocompatibility of CuNiTi arch wires<sup>(14)</sup>.

“ However, the manufacturing process of archwires containing Cu in their composition uses a sensitive and not very consistent technique”<sup>(15)</sup>

The degree,of defects in manufacturing and the composition of orthodontic applianices are factors well-known to accelerate process of corrosion and release ion by archwires <sup>(14-16)</sup> these consider that unstable “chemical structures” might cause release of ion and affect on the biocompatibility of materials<sup>(17)</sup>.

In regrard This result could also be attributed to the archwire manufacturing process. The afundmental part of biocompatibility is Resistance to corrosion and might be affected by altered factors<sup>(20)</sup> the main depends on, the type of archwire, manufactureing process and characteristics of the archwire surface.<sup>(20-23)</sup>. The studies by **Huang et al(2003)**<sup>(16)</sup> and **Huang et al(2005)**<sup>(24)</sup> decided that nickel titanium archwires of orthodontic from altered manufacturers would have resistance to corrosion in different degrees .

**Ethical Clearance:** The Research Ethical Committee at scientific research by ethical approval of both MOH and MOHSER in Iraq

**Conflict of Interest:** Non

**Funding:** Self-funding

## References

1. Nienkemper, M., Willmann, J.H., Becker, K. et al. RFA measurements of survival midpalatal orthodontic mini-implants in comparison to initial healing period. *Prog Orthod.* (2020); 21, . <https://doi.org/10.1186/s40510-020-0305-x>
2. Gurgel, J.A., Kerr, S., Powers, J.M. & LeCrone, V. (2001). Force-deflection properties of superelastic nickel-titanium archwires. *American Journal Orthodontics Dentofacial Orthopedics*,120(4): 378-82
3. Jeon, D., An, J., Lim, B. et al. Orthodontic bonding procedures significantly influence biofilm composition. *Prog Orthod.* 2020; 21, 14. <https://doi.org/10.1186/s40510-020-00314-8>
4. Shirozaki, M.U., da Silva, R.A.B., Romano, F.L. et al. Clinical, microbiological, and immunological evaluation of patients in corrective orthodontic treatment. *Prog Orthod.* (2020). 21, 6; <https://doi.org/10.1186/s40510-020-00307-7>
5. House, K., Sernetz, F. & Dymock, D. “Corrosion of orthodontic appliances—Should we care?,” *Amer. J. Orthodontics Dentofacial Orthopedics*, 2008; 133, 584–592.
6. Elayyan, F., Silikas, N. & Bearn, D. Ex vivo surface and mechanical properties of coated orthodontic archwires. *European Journal Orthodontics*, 2008; 30(6): 661-7.

7. Duffó, G.S. & Quezada, C.E. Development of an artificial saliva solution for studying the corrosion behavior of dental alloys. *Corrosion Jun*; 2004; 60(6):595-9.
8. Cawson, R.A. & Odell, E.W. *Essentials of Oral Pathology and Oral Medicine*. 7th ed., Hong Kong: Churchill Livingstone; 1998; 36-52.
9. Furlan, T.P., Barbosa, J.A., Basting, R.T., Nickel, copper, and chromium release by CuNi-titanium orthodontic archwires is dependent on the pH media. *Journal Int Oral Health*; 2018; 10, 224-228.
10. Khamees, A.M. Comparison of metal ions release and corrosion potential from different bracket archwire combinations (an in vitro study). Master thesis, College of Dentistry, University of Baghdad, Baghdad, Iraq. (2013).
11. Mohsin Sh Kh. Evaluation of corrosion pits in different types of esthetic coated orthodontic archwires in dry and wet environment at different intervals (an in vitro study). Master thesis, College of Dentistry, University of Baghdad, Baghdad, Iraq 2015.
12. Schiff, N, Lissac, M., Dalard, F., and Grosogeat, B. "FeCrNi and CoCr orthodontic brackets: evaluation of corrosion resistance in fluoridated dental rinses," *Europ. Cells Mater.*, 2005; 10, 2-7.
13. Iijima, M., Endo, K., Ohno, H., Yonekura, Y., and Mizoguchi, I. "Corrosion behavior and surface structure of orthodontic Ni-Ti alloy wires," *Dent. Mater. J.*, 2001; 20, 103-113.
14. Ramazanzadeh, B.A., Ahrari, F., Sabzevari, B. & Habibi, S. Nickel ion release from three types of nickel-titanium-based orthodontic archwires in the as-received state and after oral simulation. *J Dent Res Dent Clin Dent Prospects*. Spring; 2014; 8(2):71-76.
15. Pompei-Reynolds, E. & Kanavakis, G. Interlot variations of transition temperature range and force delivery in copper-nickel-titanium orthodontic wires. *American Journal of Orthodontics*:2014; 146 (2); 215-226.
16. Huang, H. "Variation in corrosion resistance of nickel-titanium wires from different manufacturers," *Angle Orthodontist*, 2005; 75, ( 4), 661-665 .
17. Ghazal, A. R. A., Hajeer, M. Y., Al-Sabbagh, R., Alghoraibi, I. & Aldiry, A. An evaluation of two types of nickel-titanium wires in terms of micromorphology and nickel ions' release following oral environment exposure. *Progress in orthodontics*, 2015; 16, 9-15.
18. Gopikrishnan, S., Melath, A., Ajth, V.V. & Mathews, N.B. (2015). A comparative study of bio degradation of various orthodontic arch wires: An in vitro study. *J orthodontic oral Health*;7:12-7.
19. Senkutvan, R.S., Jacob, S., Charles,A., Vadgankar, V., Jatol-Tekade, S., and Gangurde, P. Evaluation of nickel ion release from various orthodontic arch wires; An in vitro study. *Journal Int Soc Prev Community Dent*; 2014. 4:12-6.
20. Mocnik, P., Kosec, T., Kovac, J. & Biziak, M., The effect of pH, fluoride and tribocorrosion on surface properties of dental archwires. *Master Sci Eng C Mater Bio Appl*; (2017). 78:682-9.
21. Kumar, R.V., Ravlin, N., Rajakumar, P., Saravanan, R., Deepak, V.A., and Vijy Kumar, V. An accurate methodology to detect leaching of nickel and chromium ions in the initial phase of orthodontic treatment : An in vitro study. *J Contemp Dent Pract* ; (2016). 17:205-10.
22. Lages, R.B., Bridi, E.C., Perez, C.A. & Basting, R.T., Salivary level of nickel, chromium, iron, and copper in patients treated with metal or esthetic fixed orthodontics: A retrospective cohort study . *J Trace Elem Med Biol*;4(2017). 0:7-71.
23. Gravina, M.A., Canavaro, C., Elias, C.N., DAS Gracas Afonso Miranda Chves, M., Brunharo, I.H., Quintao, C.C., Mechanical properties of NiTi and CuNiTi wires used in orthodontic treatment. Part2: Microscopic surface appraisal and metallurgical characteristics. *Dental Press Journal orthodontic*; ( 2014). 19:69-76.
24. Huang, H.H., Chiu, Y. H., Lee, T. H., Wu, S. C., Yang, H. W., Su, K & Hsu, C. C. Ion release from NiTi orthodontic wires in artificial saliva with various acidities. *Biomaterials*, 2003. 24, 3585-3592.